

Defibrillator Post-use Checklist

Physio Control LIFEPAK CR Plus
Physio Control LIFEPAK CR Plus Serial Number: Defibrillator Coordinator: Defibrillator Location: <i>Complete the checklist and forward to the Occupational Health Team for retention.</i>

Refer to <i>Physio Control LIFEPAK CR Plus User Manual</i> Sections 3-5 and 5-1 to 5-8 for detailed instructions. Contact the Occupational Health Team for assistance.	Completed	Date
1. Unit and accessories are free from damage, dirt and contamination Clean and/or replace if necessary (refer to Sections 5-1 and 5-3)	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
2. Transfer data from the defibrillator, if desired (refer to Section 4-1)	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
3. New CHARGE-PAK battery charger is connected or a replacement has been ordered* (refer to Section 5-5) <i>Please note: CHARGE-PAK will usually require replacement after use</i> CHARGE-PAK battery charger has not passed expiration date	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
4. New QUIK-PAK electrode package is connected or a replacement has been ordered* (refer to Section 5-7) New QUIK-PAK electrode package has not passed expiration date	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
5. Close the lid and verify that the OK symbol appears in the readiness display, indicating the defibrillator is ready for use (refer to Section 3-5) <i>If either of the CHARGE-PAK , ATTENTION or WRENCH indicators appear, refer to the User Manual or contact authorised service provider</i>	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
Signature: (Print name if different to Defibrillator Coordinator listed above)/...../.....	

*Order replacement battery charger or electrode package immediately post use. While awaiting pad replacement, place 'OUT OF ORDER' signage on the defibrillator.